



# Aqua Solutions, Inc.

P.O. Box 70 Deer Park, TX 77536 Ph (281-479-2569)



## SPECIFICATION SHEET

### CONDUCTIVITY STANDARD 18,000 $\mu$ MHOS

Product Code	6981
Product Name	Conductivity Standard 18,000 $\mu$ MHOS
Test: QCSP30	18,000 $\mu$ MHOS @ 25° C
Prepared to Formulation on File	Yes
Appearance and Color	Clear water white liquid
Country of Origin	USA
Harmonization Code	2827.39.9000

The weights and /or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All Volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

#### PREPARATION DATE (MM/DD/YY)

Original 0.0	04/18/2018: STN
Revision	

PLEASE REFER TO THE MATERIAL SAFETY DATA SHEET FOR HEALTH, SAFETY, AND REGULATORY INFORMATION ON THIS PRODUCT.