



Aqua Solutions, Inc.

P.O. Box 70 Deer Park, TX 77536 Ph (281-479-2569)



SPECIFICATION SHEET

PHENYLARSINE OXIDE TITRANT 0.00564N

Product Code	6740
Product Name	Phenylarsine Oxide Titrant 0.00564N
APHA for Chlorine	Yes
Actual Normality	0.00564 +/- 0.00005 Normal @ 20° C
Prepared to Formulation on File	Yes
Appearance and Color	Clear water white liquid
Country of Origin	USA
Harmonization Code	2825.90.9000

The weights and /or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All Volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

PREPARATION DATE (MM/DD/YY)

Original 0.0	04/25/2018: STN
Revision	

PLEASE REFER TO THE MATERIAL SAFETY DATA SHEET FOR HEALTH, SAFETY, AND REGULATORY INFORMATION ON THIS PRODUCT.